



DS01 Dental SUCTION UNIT

OPERATION MANUAL

(Please read the operation manual before maintaining or repairing equipment, the manual is only for guidance, our company have the right to improve the design.)

After sale service: Jiangsu Dynamic Medical Technology Co.,Ltd

Thank you for choosing to use our company's DS series of medical suction machine! If the first time install and use medical suction unit, please make sure to read all the information

supplied firstly, in particular the installation information in this manual and other terms mentioned in the note. This will help you to make better use of our products.

Disclaimer

The description of this manual does not mean any description of Dynamic product specification and configuration. For the product specifications and configuration, please refer to the related products with Dynamic manual, packing list, or directly consult the product dealer.

In the preparation for manual process, has sought to correct and complete content, but because of technical and technological advances, relevant laws and regulations changed, our medical suction machine and operation manual will be changed accordingly. Therefore, we have the right to revise this manual description of product and the contents of this manual at any time, without notice. Dynamic will continue to constantly improve their products and services.

If you find the actual situation of the products and the instructions are inconsistent when using the operation manual, or you want to get the latest information, or have any questions, please feel free to tell us.

Limited warranty

If the product comes with the details of our service (warranty documents), we will provide you the service according to this detail service. If the product does not come with our company's service rules (warranty file), it indicates that your after-sales service will be executed according to the written agreement of the product provider and the product receiver, it is not suitable for service conditions (warranty documents), and it do not contain any ostensive or implicit warranty, in this situation, you can not get the our after-sales service, if you are willing to accept the our product, you should require the organizations or persons who provides you the products to provide you with their service according to their commitment.

These products are only in accordance with local laws and regulations of the national and regional sales.

The maximum extent permitted by law, we will not bear any responsibility for the company in the following cases:

- 1, Third-party claims against you;
- 2, Incidental or indirect damages and economic losses;
- 3, Due to your own configuration and all damages caused by the economic loss;
- 4, Because you are not required in the environment in this guide use of this product, or any damage and economic loss caused by the failure of this method of operation under the guidance;
- 5, all damage and economic loss caused by force Majeure.

You can call the sales person.

Note: The phone may be subject to change because of the communications network or other objective factors without notice.

Copyright

This manual was owned and all rights were reserved by our company. Other accessories of the product or product documentation, the copyright belongs to the corresponding organization or person owned.

Sign flag



Note □ There are likely to lose parts and material document



Note □ That can damage equipment or cause equipment not working properly

Marginal data



ult Instructions for use



Protective Earth



On



ate of Manufacturer



Off



Manufacturer

Contents

- 1. Brief introduction**
- 2. Technical Parameter**
- 3. The Equipment Structure and Installation**
- 4. Functions of main parts**
- 5. Maintenance**
- 6. Split-up drawing of ring blower**
- 7. Troubleshooting**
- 8. Working principle overview**
- 9. Safety precautions**
- 10. Drawing of air connection and electrical circuit**
- 11. External mark and content**
- 12. Packing list**

1. Brief introduction

Thank you for purchasing our dental suction unit. It is our company well-developed multi-purpose mechanical and electrical integration of energy-saving product with small, mobile, good appearance, less power consumption, free maintenance, safe and durable etc. In order to be convenient for you to operate the unit, please read the manual carefully.

1.1 Contents of the manual

The operation manual includes brief introduction, technical data, installation and test, maintenance, split-up drawing of ring blower and so on. You can get help in according to the contents.

1.2 Performance and structure

Medical suction unit consists of air and water separator, filter, ring blower, draining pump and electrical board.

1.3 Usage

The suction unit is applied in the process of oral cavity treatment as it can pump foreign objects attached with germina, such as oral secretion and dirty blood, as well as protecting environment since the objects can be dealt with before being pumped. It is mainly used for clinics and hospitals in the treatment of oral health.

1.4 Safety Information

1.4.1 Electric safety

- To make sure that the plug of this unit is three-phase with good grounding.
- Before connecting the power plug, please ensure the rated voltage consistent with the named voltage which this unit required.
- To avoid unsteady voltage, please do not use the power plug together with other electric equipment.
- Before maintaining or repairing the unit, please unplug that electric wire firstly.
- Please check the electric wire and plug regularly to avoid the damage or extrusions.

1.4.2 Cleaning

Please keep the environment around the product clean, and note to unplug that electric wire before cleaning the unit. The process of cleaning as follows: firstly wipe the outer surface with a soft cloth stained with neutral detergent, and then use a soft cloth to clean the machine again.

Notice: Do not use liquid or detergent containing flammable substances.

1.4.3 Other safety information

When using this unit, please strictly operate it in according to the following rules:

1) Please note that the power plug is designed as the safety device, not power switch. It is necessary to ensure the power plug can be unplugged from the plug without extrusions.

Warning: The good grounding is necessary to avoid the risk of electric shock.

2) Equipment cannot be used in flammable mixture environment.

3) Equipment cannot be used in anesthetic and air, oxygen, nitrogen mixture environment.

4) Don't put the equipment together with oxygen cylinder.

5) Please do not operate the equipment or replace the accessories when the following situations occur:

- Damage of the power plug
- Equipment cannot work normally
- Breakdown of the equipment
- Water inlet of the equipment
- Loud or shrill noise, overheated outlet air or foul smells during operation.

When above situations occur, please contact the manufacturer or local agent. In order to be convenient for us to offer after sale service, please offer drawing of electrical circuit, air and water connection diagram, packing list and other useful information when necessary.

6) Please use original or qualified disinfectant.

7) Turn off the power when the operator left.

8) Dispose of waste liquid and waste solid according to the local health regulations.

9) Do not use and store the equipment outside the specified environment.

10) Keep accessories from dropping to avoid damages.

11) Keep the electric wire from loosening in the process of treatment.

12) Please use accessories from original manufacturer.

13) Keep the equipment in a solid and horizontal level.

1.5 Product information

Name: Medical Suction Unit

Model: DS

Configuration: refer to enclosed Packing List

Models of dental suction units: DS01

2. Technical data

2.1 Specifications:

Sheet 1

Product Model	Volt/Freq	Power□VA)	Suction Air Flow□L/ min□	Max. Negative Pressure (kPa)	Noise Level(dB)	Number of Chair
DS01	AC230V 50Hz	370	300	-11	≤65	1/2
	AC115V 60Hz	400	300	-11	≤65	1/2
	AC220V 60Hz	400	300	-11	≤65	1/2

2.2 Configurations

Sheet 2

Product Model	Ring Blower	Filter	Air and Water Separator	Control Board	Amalgam Separator
DS01	●	●	●	●	○

Note: ● means Standard, ○ means Optional



Warning:

Please operate the unit strictly under the rated supply voltage, because the unsteady current can cause injuries to the unit.

Please install [power voltage regulator](#) when supply voltage is unsteady, and the rated power is no less than 4000VA.

2.2 Product Classification

2.2.1 Classification of air flow: High volume suction system

2.2.2 Classification of suction method: semi-dry system

2.2.3 Classification of dental equipments management: Oral equipments and instruments

(code6855), Type 1 management

2.2.4 Classification of electric shock protection: Type I equipment.

2.2.5 Degree of Protection against electric shock: Type B equipment.

2.2.6 Classification of waterproof: Normal equipment (IP20).

2.2.7 Classification of working frequency: intermittent duty.

2.3 Transportation, storage and condition of usage

2.3.1 Condition of transportation

The machine should be transported according to specific requirements of signed contract.

2.3.2 Condition of storage

The machine should be stored in the place with relative humidity not greater than 80%. Surrounding area of the machine should be clean, dry, and free of corrective gas, well ventilated and indirect sunlight.

2.3.2 Condition of usage

Ambient temperature: 5°C~40°C

Relative humidity: ≤80%

Atmosphere pressure: 86kPa ~106kPa

Warning: The unit can not be used in an area where the air is contaminated with the flammable [anaesthetic](#), which the explosion may occur.

2.4 Condition of installation

Installing the machine in a solid and horizontal level with, and does not have an incline greater than 2°.

2.5 Condition of power

Voltage range: AC110-220V 50/60Hz

2.6 Electromagnetic Compatibility

Electromagnetic Emission Standards and Our Announcement		
DS series dental suction unit should be used in following electromagnetic environment, and the use of this product should ensure to use under suitable environment.		
Electromagnetic Emission Test	Standards	Electromagnetic Environment Standards
Radiation Electromagnetic Disturbance	CISPR 11 I FORM A CLASS	Electromagnetic emission of DS series dental suction unit can only satisfy its inherent functions, so its wireless radiation level is so low, that could not effect the contiguous electronic equipment.
Power Disturbance Voltage	CISPR 11 I FORM A CLASS	
Harmonic Emission	IEC 61000-3-2	DS series dental suction unit can used by all companies and institutes, and could not connect with household or public low voltage power system.
Voltage Fluctuation/Flash	IEC 61000-3-3	

Electromagnetic Emission Standards and Our Announcement			
DS series dental suction unit should be used in following electromagnetic environment, and the use of this product should ensure to use under suitable environment.			
Electromagnetic Emission Test	IEC 60601 TEST VALUE	Standards	Electromagnetic Environment Standards
Electrostatic Discharge (ESD) IEC 61000-4-2	±6 kV Touch ±8 kV Air	±6 kV Touch ±8 kV Air	The materials of floor should be wood, concrete and ceramic tile. If the floor is synthetic material, relative humidity no less than 30%.
EFT IEC 61000-4-4	±2kV Power Supply Line	±2kV Power Supply Line	The quality of power supply is used in special commerce or hospital.
Surge (Impact) IEC 61000-4-5	±1 kV Line to Line Mode ±2 kV Line to Earth Mode	±1 kV Line to Line Mode ±2 kV Line to Earth Mode	The quality of power supply is used in special commerce or hospital.
Voltage Sag And Short	<5% UT	<5% UT	The quality of power supply is used in special

Supply Interruption Voltage Change IEC 61000-4-11	(>95% dip in UT) for 0.5 cycle 40% UT (60% dip in UT) for 5 cycles 70% UT (30% dip in UT) for 25 cycles <5% UT (>95% dip in UT) for 5 sec	(>95% dip in UT) for 0.5 cycle 40% UT (60% dip in UT) for 5 cycles 70% UT (30% dip in UT) for 25 cycles <5% UT (>95% dip in UT) for 5 sec	commerce or hospital. When DS series dental suction unit need to continuous operation when grid interrupts, please use uninterruptable power output.
Power Frequency Magnetic Fields IEC 61000-4-8	3A/m	3A/m	/
Note: UT is alternating voltage before testing.			

Electromagnetic Emission Standards and Our Announcement			
DS series dental suction unit should be used in following electromagnetic environment, and the use of this product should ensure to use under suitable environment.			
Protection Test	IEC 60601 TEST VALUE	Standards	Electromagnetic Environment Standards
Radio transmission IEC 61000-4-6 Radio radiation IEC 61000-4-3	3 Vrms 150 kHz to 80 MHz 3 V/m 80 MHz to 2.5 GHz	3 Vrms 3 V/m	Portable or mobile radio communication equipment should stay away from DU series portable dental unit including cables, also should leave more than the recommended separation distance calculated. Recommended separation distance $d \geq \frac{3.5}{V_1} \sqrt{P}$ $d \geq \frac{3.5}{E_1} \sqrt{P} \quad 80 \text{ MHz } 800 \text{ MHz}$ $d \geq \frac{7}{E_1} \sqrt{P} \quad 800 \text{ MHz } 2.5 \text{ GHz}$ P: is the fixed output power by radio equipment manufacturer, the unit is W. D: Recommend separation distance. Fixed radio transmitter, determined by electromagnetic measurements, A: Each frequency band must be less than the standard requirements. B: The equipment maybe interfered nearby the device with following symbols: 
Note: 1 Between 80 MHz and 800 MHz, higher frequency range applies. 2 This specification maybe do not suitable for all situations, because the propagation of electromagnetic waves can be effected by the absorption and reflection of buildings, objects and people. A: The electromagnetic strength of fixed radio transmitter, such as radar station, cordless telephone and mobile radios, amateur radio, FM and AM radio and television, this can not be predicted in theory. The estimated of electromagnetic environment depends on the fixed radio transmitters, and electromagnetic measurement points should be considered. If the magnetic strength nearby the DU series is more than the above allowed value,			

then DU series portable dental unit should be observed to verify by the normal operation. In addition, when to re-install the DU series portable dental unit, the above information should you measured again.

B: Frequencies above 150kHz to 80 MHz, the electromagnetic intensity should be less than 3V/m.

Recommended separation distance between DS series dental suction unit and Portable or mobile radio communication equipment

DS series dental suction unit for radio interference to be controlled electromagnetic environment, the users of DS series dental suction unit should ensure the minimum distance between the DS series dental suction unit and portable or mobile radio communication equipment according to the following recommended value, and take the maximum output power into account.

Communications equipment, the rated maximum output power(W)	Separate communication channels established in accordance with the distance (m)		
	150 kHz to 80MHz	80 MHz to 800 MHz	800 MHz to 2.5GHz
	$d \geq \frac{3.5}{V_1} \sqrt{P}$	$d \geq \frac{3.5}{E_1} \sqrt{P}$	$d \geq \frac{7}{E} \sqrt{P}$
0.01	0.117	0.117	0.233
0.1	0.369	0.369	0.738
1	1.167	1.167	2.333
10	3.689	3.689	7.379
100	11.667	11.667	23.333


Because the rated maximum output power of communications equipment do not mentioned above, the recommended separation distance d (meters) can be estimated according to same communication device. P is the maximum output power (in watts) according to the communications equipment manufacturers.

Note: 1. From 80 MHz to 800 MHz, separation distance is more accordance with the higher frequency.

2. This specification do not apply to all situations. Because the propagation of electromagnetic waves by buildings, objects, people effects of absorption and reflection.

3. The Equipment Structure

3.1. structure:DS01



1. Air Pump

2.Filter Pipe

3. Power Switch

4. Overload Protector	5. Decompression Valve	6. Air Outlet
7. Suction Port	8. Drain Outlet	

3.2 Procedures of installation

3.2.1 Installation:

□1□ Check after unpacking

Ensure that every component of the unit is intact after opening the box and check that spare parts are intact according to packing list. In case of doubt, please contact our company or local agent.

□2□ Precautions of installation

The electric units should be installed by competent person only.

□3□ Condition of installation

(a) Ambient temperature: 5°C~40°C

(b) The unit should be well positioned to ensure there is an adequate airflow around the machine.

□4□ Method of installation

(a) Install the equipment close to treatment unit, and ensure mounting level of suction unit lower than treatment units. Please note that the door of suction unit should not be opened freely for safety.

(b) The suction unit should be well situated to ensure it has 25cm space free of obstacle around two sides of case.

□5□ Specifications of connecting pipes

(a) Pipe connects suction unit with dental treatment unit□

Material: PVC or PPR

Inner diameter: φ32

Length: ≤20m

Height of installation: 0.5m higher than bottom of suction unit

(b) Pipe connects air outlet with outdoor pipeline:

Material: PVC or PPR

Inner diameter: φ32

(c) Pipe connects drain water exit with outdoor waste processing pipeline:

Material: PVC or PPR

Inner diameter: φ32

□6□ Electrical requirements

Power: when install the power socket; please ensure the power voltage consistent with the rated voltage and frequency which this unit required.

3.2.2□ Testing

Check all connections of tubes, voltage and current, and connect drain outlet port into drain pipe. After connected choose valve and signal lines, open power switch, and indicator light of power switch is on. Lift up saliva ejector from rack, and operation of suction unit starts. Water will be sucked by suction port, and pump out from drain port, then put saliva ejector to original place of rack. Suction unit will stop after delay of 10 seconds.



Warning: To avoid unsteady voltage, please do not use the power plug together with other electric equipment.

4. Functions of main parts

Filter: mainly used to collect and separate solid particles (such as tooth chippings, amalgam particles), bloods and secretions, then be more convenient for centralized processing and protect environment.

Note: please clean filter weekly.

Ring Blower: mainly offer negative pressure for dental treatment.

Air and water separator: ring blower offers negative pressure to pump drainage water into air and water separator, then drainage water and air will be separated individually, and achieve function of pumping in and out at same time.


Amalgam separator: optional function, separation rate must be not less than 95%, and metal collection bottle assembled with amalgam separator will collect these heavy metals to avoid pollutions to environment.

5. Maintenance

1. Clean the suction tubes everyday by sucking clean water twice (1 liter each time) after use.
2. Clean the drain bottle and filter weekly.
3. Clean the tubes every three months.

6. Split-up drawing of ring blower

001 Wind cover	006 Bushing block	011 Bearing	016 Connector
002 Screw	007 Wind barrel	012 Electric motor	017 Foot
003 Spring-pad	008 Bearing	013 Rubber gasket	
004 Impeller plate	009 Bearing block	014 Core	
005 Impeller	010 Rotor	015 Flue pipe	



7. Troubleshooting

	Problem	Reason	Check
1	Cannot work with very weak suction	The air inlet is blocked	Check the air inlet
		The filter is blocked	Check the filter
2	Can not start up the compressor with the compressor shaking and noisy	The power pressure is too low.	Make the pressure rise or use manostat.
3	The pump cannot work.	The power is not switched on	Check if the power supply is connected correctly
		A fuse is blown	Check if the fuse is blown
		The overload protector is bounced	Check if the overload protector is bounced
		The wire to the power supply has fallen off or an electronic component is loose	Check the patch board and the electric connection of the compressor

8. Working principle overview

Ring blower working principle: Ring blower's impeller contains numerous blades in its edge and it is special designed for a high air pressure. When the impeller is rotating, air between the impeller would be accelerated along the direction of the fan blades and due to rotating movement, after air enters outside air layer, it comes back to the base of the fan blades due to the effect of the differential pressure, the cycle will repeat several times to make the air run spirally so as to increase the pressure. The air will expel quickly because of the pressure there bigger than the pressure in the system when the air runs through the outlet.


Operating principle: Connect the suction unit with dental treatment unit, take up saliva ejector from clip-on instrument holder of dental treatment unit, and the suction unit will start automatically to suck foreign objects attached with germina (such as oral secretion and dirty blood) into filter, the filter will separate solid particles(such as teeth fragments, silver amalgam particles) from dirty blood and oral secretion(please clean solid particles inside the filter regularly).The filtered dirty blood and oral secretion will be sucked into drain bottle, the air and water separator will separate the air from dirty blood and oral secretion to ensure the air offered for ring blower is clean and prevent the impellers from corrosive damage.

9. Safety precautions


1. When dental suction units are working, please make sure that there will be no goods can be touched the air compressor.

2. Clean the filter regularly.
 3. To change the electrical parts timely, but must turn off the power.
 4. When to clean and maintenance the dental units, please to turn off the power.
 5. After setting up the suction units, do not move the machine too often, avoid to affecting its working conditions.
 6. To forbid the unprofessional person to use the air compressor, avoiding to operating wrong.
 7. When to maintenance the suction units, it must be some person who is very qualified.
 8. Do not let the olds, children come into contact with the compressor to avoid resulting harm.
 9. When to use the suction units under knowing there will be result of harm, to forbid using the suction units.
 10. When the dental suction units are retired, the capacitor and electrical components should be available to the local law and regulation.
- 10. Drawing of air connection and electrical circuit**

10.1 DS01 air connection



10.2 DS01 electrical schematic diagram



11. External mark and content

1. Following are concluded in external packing:

Product name: Medical Suction Unit

2. Model: DS

3. Standard NO.:

4. Product registration NO.:

5. Factory name: Jiangsu Dynamic Medical Technology Co.,Ltd

Factory address: No.108 Xingpu Road, Lujia Town, Kunshan, City, Jiangsu, 215331

PEOPLE'S REPUBLIC OF CHINA


6. Packing dimension:

External size: L×W×H

G.W: kg

N.W: kg

7. Graph label



12. Packing List

No.	Name	Quantity
1.	DS01 Suction Unit	1
2.	ID Φ32 Tubes	1
3.	ID Φ20 Tubes	1
4.	Tube Hoop Φ40	6
5.	ID 20mm Tube(1.5m)	1
6.	Tube Hoop Φ20	2
7.	Electrical Wires	1
8.	Operation Manual	1
9.	Warranty Card	1
10.	Qualification Certificate	1